## VivaScope

## Medical > In Vivo > Ophthamology

## Handheld Reflectance Confocal Microscopy for the Diagnosis of Conjunctival Tumors.

*Cinotti E, Perrot JL, Labeille B, Campolmi N, Espinasse M, Grivet D, Thuret G, Gain P, Douchet C, Andrea C, Haouas M, Cambazard F., Am J Ophthalmol. 2014 Nov 5. pii: S0002-9394(14)00696-5. doi: 10.1016/j.ajo.2014.10.031.* 

## ABSTRACT

**PURPOSE:** To evaluate whether the handheld in vivo reflectance confocal microscopy that has been recently developed for the study of skin tumors is suitable for the diagnosis of conjunctival tumors. **DESIGN:** Prospective study, observational case series.

**METHODS:** We prospectively evaluated the reflectance confocal microscopy features of 53 conjunctival lesions clinically suspicious for tumors of 46 patients referred to the University Hospital of Saint-Etienne (France) by using the handheld device. Twenty-three lesions were excised (3 nevi, 10 melanomas, 5 squamous cell carcinoma, 2 lymphomas, and 3 pinguecula/pterygium) while the other 30, presenting no reflectance confocal microscopy malignant features, were under follow-up for at least 1 year. Clinical reflectance confocal microscopy and histologic diagnosis were compared.

**RESULTS:** In vivo reflectance confocal microscopy diagnosis was in agreement with the histologic diagnosis in all cases and none of the lesions that were not excised show any clinical progression under follow-up. **CONCLUSION:** In vivo reflectance confocal microscopy with a handheld

dermatology-dedicated microscope can play a role in the noninvasive diagnosis of conjunctival lesions. Further studies should be performed to better define the diagnostic ability of this technique.